


Problem with the surface reference sample

Test the surface reference sample in order to check if it had any damage


Diameter: 10 mm
Step: 0.1mm



- AC signal (expected to be proportional to the absorption), has periodic oscillations by a factor of ~ 2
- DC and phase are pretty constant but follow the same pattern
- Looks like real absorption
- Looks like the sample is damaged and absorption is not uniform


Made a smaller and more resolved map around the center
Then rotated the sample by 90° and the pattern rotates.

Diameter: 2mm
Step: 0.05mm




Rotated by 90°

Diameter: 2mm
Step: 0.05mm




- Looks like the problem is in the sample

- Flipped the sample and made a new scan




Z=34.7mm



Z=30.9mm



- Moved the sample to the new maximum position and made a map



The pattern is a bit different and with lower amplitude

Check if the same pattern is in the **bulk reference** sample


Diameter: 2mm
Step: 0.05mm




THE PATTERN DOESN'T SHOW IN THE BULK REFERENCE SAMPLE

Bought a **NEW surface reference** sample:
Optical density 0.2 filter from Newport (model: FRQ-ND02)

Diameter: 2mm
Step: 0.05mm



Diameter: 10mm
Step: 0.1mm




Also the new sample shows a similar pattern but with larger fringes separation:

- **New sample: 5 fringes/cm**
- **Old sampe: 25 fringes/cm**

NEW surface reference sample

Map along Y and Z axis



This map shows the cross section of the coating with the absorption fringes (like many scans along Z for different Y positions)

- The damage reason is excluded by the fact that we can see a similar pattern on the new sample.
- Another possible reason is the wedge of the sample, which is different for different samples, in combination with a misalignment. Although the specs say: wedge < 3 arc min
- I will try to make a better alignment of the crossing point and see if something changes